

2023-05-31

Anti-hCG 5016 SPRN-5

Product overview

100013 Catalog number

Specificity Antibody recognizes human chorionic gonadotropin and its free beta

subunit

Description Monoclonal mouse antibody, cultured in vitro under conditions free from

animal-derived components.

Product buffer solution 37 mM citrate, 125 mM phosphate, pH 6.0, 0.9 % NaCl, 0.095 % NaN₃ as

a preservative

Shelf life and storage 36 months from manufacturing at 2-8 °C

Subclass IgG₁

Analyte description Human chorionic gonadotropin (hCG) is a glycoprotein hormone produced

in pregnancy by the developing embryo soon after conception and later by the syncytiotrophoblast (part of the placenta). Its role is to prevent the disintegration of the corpus luteum of the ovary and thereby maintain progesterone production that is critical for a pregnancy in humans. Early pregnancy testing, in general, is based on the detection of hCG. hCG is produced also by some tumors, but it is not known whether this production

is a contributing cause or an effect of tumorigenesis.

Parameters tested on each lot

Product appearance Liquid, may turn slightly opaque during storage

5.0 mg/ml (+/- 10 %) **Product concentration**

Immunoreactivity 80-120 % compared to the reference sample in an FIA test

IEF Profile 6.2 - 6.8

Purity ≥ 95 %

Oy Medix Biochemica Ab

FI-02180 Espoo, Finland

Klovinpellontie 3,

Kinetic parameters

hCG: 1.5×10^6 1/Ms and hCG β : 1.9×10^6 1/Ms **Association rate constant**

Dissociation rate constant hCG: 9.5×10^{-4} 1/s and hCG β : 5.2×10^{-3} 1/s

Affinity constant hCG: $K_A = 1.6 \times 10^9 \text{ 1/M}$; $K_D = 6.4 \times 10^{-10} \text{ M}$ (= 0.64 nM)

hCG β : K_A = 3.5 x 10⁸ 1/M; K_D = 2.8 x 10⁻⁹ M (= 2.8 nM)

Determination method SPR analysis (ProteOn XPR36)

Determination antigen hCG, Scripps (Cat C0714); hCGβ, Scripps (Cat C0914)





Product specification ANTIBODY

2023-05-31

Cross-reactivities Does not recognize hCGa, LH, FSH, or TSH

Epitope Beta-2 as described in Berger et al. (2013). The antibody recognizes both

intact hCG and free β subunit.

Pair recommendations

		DETECTION										
		hCG beta								alpha subunit		
		5004	5006	5008	5009	5011	5012 free β	5014	5016	5501	5503	6601
CAPTURE	5004	-	-	-	+	+	-	+	-	+	+	+
	5006	-	-	-	-	-	-	+	-	+	+	+
	5008	-	-	-	+	-	-	+	-	+	+	+
	5009	+	+	+	-	-	-	+	+	-	-	+
	5011	+	+	+	-	-	-	+	+	-	-	+
	5012 free β	+	+	+	-	-	-	+	+	-	-	-
	5014	+	+	+	+	+	-	-	+	+	+	+
	5016	-	-	-	-	-	-	+	-	+	+	+

Following pairs are especially recommended for free hCG beta assays: CLIA: 5012 (capture) - 5004 (detection) and 5012 - 5008

Please note that pair recommendations are based on results obtained by our laboratory. Equally good results may be obtained using other pairs and therefore these recommendations are only indicative.

Platforms tested FIA, CLIA

Antigens tested Native hCG antigens, Medix Biochemica 189-10 and 189-11

Native β-hCG antigen, Medix Biochemica 325-11

Product stability TEMPERATURE, TIME RESULT

-70 °C, 21 days Not Determined (N/D)

-20 °C, 21 days OK +4 °C, 21 days OK +35 °C, 21 days OK +45 °C, 7 days OK

Stability testing is performed in the product buffer to see whether different temperatures affect the antigen binding, charge or composition of the antibody. Please note that the shelf life given on the first page is based on real time stability testing at 2-8 °C in the product buffer.

Miscellaneous





Product specification ANTIBODY

2023-05-31

References

Berger, P., Paus, E., Hemken, P.M., Sturgeon, C., Stewart, W.W., Skinner, J.P., Harwick, L.C., Saldana, S.C., Ramsay, C.S., Rupprecht, K.R., Olsen, K.H., Bidart, J.M. and Stenman, U.H. (2013) Candidate epitopes for measurement of hCG and related molecules: the second ISOBM TD-7 workshop. Tumor Biol., 34: 4033-4057.

